

Instrument Installation Report

Instrument Name sn12c020920
Serial Number
Chip Surface
Initial pH: 5.01 Thu Dec 19 04:10:56 2024

1) W2 pH=5.01
1) Adding 24.70 ml
2) W2 pH=5.22
2) Adding 25.00 ml
3) W2 pH=5.46
3) Adding 25.00 ml
4) W2 pH=5.71
4) Adding 25.00 ml
5) W2 pH=5.97
5) Adding 25.00 ml
6) W2 pH=6.25
6) Adding 25.00 ml
7) W2 pH=6.55
7) Adding 25.00 ml
8) W2 pH=6.98
1) W2 pH=6.99
1) Adding 5.08 ml
2) W2 pH=7.12
2) Adding 11.14 ml
3) W2 pH=7.40
3) Adding 2.38 ml
4) W2 pH=7.46

Total Hydroxide Added: 193.30 ml
Total Iterations: 12
Final pH: 7.46 Thu Dec 19 05:11:57 2024

Raw Traces Shift=0.33
RawTraces R4: 7.11
RawTraces R3: 7.09
RawTraces R2: 7.02
RawTraces R1: 7.08
RawTraces W2: 7.46
RawTraces W1: 10.09

Instrument Installation Status: **PASSED**

Acknowledged by:

Customer Signature

Date

Life Tech FSE Signature

Date